

Safe-T-Vue® 6 • Product No. 7206

Quality Assurance: Safe-T-Vue® 6 Non-Reversible Blood Bag Temperature Indicator Product No. 7206
 PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE® 6 IS PRINTED IN GREEN (PMS 327).

Test performed on Random Sampling of Finished Product **Lot # S6 102612**
 Expiration Date: 2014-10

Non - Chemistry

Mark \checkmark for "accepted" and X for "Not Accepted".

Box	Adh. Label	Foil Lid	Close Function	Fill Level
\checkmark Copy	\checkmark Copy	\checkmark Copy	\checkmark Formed	\checkmark Acceptable
\checkmark Tack	\checkmark Tack	\checkmark Position	\checkmark Locked/Stable	Underfill
\checkmark Clean	\checkmark Color	\checkmark Clean Peel	Hinge	Overfill
\checkmark Box Damage	Tray	Disk	\checkmark Elasticity	Color Label
	\checkmark Edges	\checkmark Appearance		\checkmark Color
	\checkmark Color	\checkmark Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003

Performed by using a "DigiSense" Thermister Electronic Thermometer. Using a cold (1° to 4° C) bag of simulated blood (a glycerol-water-red dye solution) with the electric thermometer placed and sealed into the core of the blood bag to permit observation of temperature changes. For detailed protocol request QAP 0300-01 REV 0 Mar 01, 2005 [Incoming inspection of 7208 follows the same protocol for Safe-T-Vue® 10]

Temperature Change and Corresponding Safe-T-Vue® Change

Sample Set 1	DATA	Sample Set 2
1 to 4 °C	<i>Starting Temperatures</i>	1 to 4 °C

OBSERVATIONS

Temperature	Condition
5.0 °C	Red Spots - Random WHITE CENTER
5.5 °C	Red Filling In WHITE CENTER
5.9 °C	Full Red Color RED CENTER

Temperature	Condition
5.0 °C	Red Spots - Random WHITE CENTER
5.5 °C	Red Filling In WHITE CENTER
5.9 °C	Full Red Color RED CENTER

Temperature accuracy is +/-0.3°C.

Results of Testing Lot # **S6 102612**; Test Date: Nov 8-12, 2012

Red Spots - Random 5.0 C; Red Filling in White Center 5.5 °C; Full Red Color at 5.9°C.

