



Safe-T-Vue® • Product No. 7201

QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Non-Reversible Temperature Indicator Product No. 7201

Lot # **S10 051315**

Expiration Date: May 2017

Non - Chemistry

Mark \checkmark for "Accepted" and X for "Not Accepted."

| Box | Adh. Label | Foil Lid | Close Function | Fill Level |
|-------------------------|--------------------|-------------------------|----------------------------|-------------------------|
| \checkmark Copy | \checkmark Copy | \checkmark Copy | \checkmark Formed | \checkmark Acceptable |
| \checkmark Tack | \checkmark Tack | \checkmark Position | \checkmark Locked/Stable | Underfill |
| \checkmark Clean | \checkmark Color | \checkmark Clean Peel | Hinge | Overfill |
| \checkmark Box Damage | Tray | Disk | \checkmark Elasticity | Color Label |
| | \checkmark Edges | \checkmark Appearance | | \checkmark Color |
| | \checkmark Color | \checkmark Fit | | |

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.

| Temperature Change and Corresponding Safe-T-Vue® Change | |
|---|--|
| Product Specification | Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C |
| Starting Test Temperatures | 1 to 4 °C |

| Test Parameter | AQL | Sample Size | Accept Number | Test Result | Pass/Fail |
|---------------------------------|-----|-------------|---------------|-------------|-----------|
| Full Color Response Temperature | 1% | 125 | 2 | Pass | Pass |

Quantity Approved: 60,600

Performance Data: Red Spots - Random 9.7 ° C; Red Filling in White Center 9.9 ° C; Full Red Color at 10.0 ° C.

Quality Control Verification: [Signature] 03 Aug 15
Quality manager Date

Quality Assurance: Approved Lot # S10 051315; [Signature] 3 Aug 15
Quality manager Date