

Safe-T-Vue® • Product No. 7206
QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® X Non-Reversible Temperature Indicator Product No. 7206

 Lot # **S6 013014**

 Expiration Date: JAN-16
Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 3.5 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	80	28350	Pass	Pass

Performance Data: Red Spots - Random 5.1 °C; Red Filling in White Center 5.5 °C; Full Red Color at 6.0 °C.

 QUALITY ASSURANCE: APPROVED LOT # S6013014

QAP/0102-2, REV 010, 15JAN14


 Quality Manager

25 Feb 14
 Date