

**Safe-T-Vue® 6 • Product No. 7206**

**Quality Assurance:** Safe-T-Vue® 6 Non-Reversible Blood Bag Temperature Indicator Product No. 7206  
 PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE® 6 IS PRINTED IN GREEN (PMS 327).

Test performed on Random Sampling of Finished Product **Lot # S6 060712**  
 Expiration Date: 2014-06

**Non - Chemistry**

Mark ✓ for "accepted" and X for "Not Accepted".

Box	Adh. Label	Foil Lid	Close Function	Fill Level
✓ Copy	✓ Copy	✓ Copy	✓ Formed	✓ Acceptable
✓ Tack	✓ Tack	✓ Position	✓ Locked/Stable	Underfill
✓ Clean	✓ Color	✓ Clean Peel	<b>Hinge</b>	Overfill
✓ Box Damage	<b>Tray</b>	<b>Disk</b>	✓ Elasticity	<b>Color Label</b>
	✓ Edges	✓ Appearance		✓ Color
	✓ Color	✓ Fit		

**Test Protocol – Finished Goods Quality Assurance**

*Test performed on random sampling of finished product per ANSI ASQZ1.4-2003*

Performed by using a "DigiSense" Thermister Electronic Thermometer. Using a cold (1° to 4° C) bag of simulated blood (a glycerol-water-red dye solution) with the electric thermometer placed and sealed into the core of the blood bag to permit observation of temperature changes. For detailed protocol request QAP 0300-01 REV 0 Mar 01, 2005 [Incoming inspection of 7208 follows the same protocol for Safe-T-Vue® 10]

**Temperature Change and Corresponding Safe-T-Vue® Change**

Sample Set 1	DATA	Sample Set 2
1 to 4 °C	Starting Temperatures	1 to 4 °C

**OBSERVATIONS**

Temperature	Condition
4.3 °C	Red Spots - Random WHITE CENTER
5.2 °C	Red Filling In WHITE CENTER
5.9 °C	Full Red Color RED CENTER

Temperature	Condition
4.2 °C	Red Spots - Random WHITE CENTER
5.2 °C	Red Filling In WHITE CENTER
5.8 °C	Full Red Color RED CENTER

Temperature accuracy is +/-0.3°C.

**Results of Testing** Lot # S6 060712; Test Date: June 26, 2012

Red Spots - Random 4.3 C; Red Filling in White Center 5.2 ° C; Full Red Color at 5.9° C.

