



**Safe-T-Vue® • Product No. 7201**

**QUALITY ASSURANCE LOT PERFORMANCE REPORT**

Safe-T-Vue® 10 Non-Reversible Temperature Indicator Product No. 7201

Lot # **S10 091815**

Expiration Date: Sep 17

**Non - Chemistry**

Mark ✓ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
✓ Copy	✓ Copy	✓ Copy	✓ Formed	✓ Acceptable
✓ Tack	✓ Tack	✓ Position	✓ Locked/Stable	Underfill
✓ Clean	✓ Color	✓ Clean Peel	<b>Hinge</b>	Overfill
✓ Box Damage	✓ Tray	Disk	✓ Elasticity	<b>Color Label</b>
	✓ Edges	✓ Appearance		✓ Color
	Color	✓ Fit		

**Test Protocol – Finished Goods Quality Assurance**

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from [www.williamlabs.com](http://www.williamlabs.com).

<b>Temperature Change and Corresponding Safe-T-Vue® Change</b>	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	Pass	Pass

Quantity Approved: 60,700

**Performance Data:** Red Spots - Random 9.6 ° C; Red Filling in White Center 9.9 ° C; Full Red Color at 10.1 ° C.

Quality Control Verification: [Signature] 11 Nov 15  
Quality manager Date

Quality Assurance: Approved Lot # S10091815; [Signature] 11 Nov 15