

**Safe-T-Vue® • Product No. 7201**  
**QUALITY ASSURANCE LOT PERFORMANCE REPORT**  
 Safe-T-Vue® **10** Non-Reversible Temperature Indicator Product No. **7201**

Lot # **S10 070716**

Expiration Date: **2018-07-07**

**Non - Chemistry**

Mark  $\checkmark$  for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
$\checkmark$ Copy	$\checkmark$ Copy	$\checkmark$ Copy	$\checkmark$ Formed	$\checkmark$ Acceptable
$\checkmark$ Tack	$\checkmark$ Tack	$\checkmark$ Position	$\checkmark$ Locked/Stable	Underfill
$\checkmark$ Clean	$\checkmark$ Color	$\checkmark$ Clean Peel	<b>Hinge</b>	Overfill
$\checkmark$ Box Damage	$\checkmark$ <b>Tray</b>	<b>Disk</b>	$\checkmark$ Elasticity	<b>Color Label</b>
	$\checkmark$ Edges	$\checkmark$ Appearance		$\checkmark$ Color
	$\checkmark$ Color	$\checkmark$ Fit		

**Test Protocol – Finished Goods Quality Assurance**

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from [www.williamlabs.com](http://www.williamlabs.com).

<b>Temperature Change and Corresponding Safe-T-Vue® Change</b>	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	PASS	PASS

Quantity Approved: \_\_\_\_\_ 59,950 \_\_\_\_\_

**Performance Data:** Red Spots - Random **9.7 °C**; Red Filling in White Center **9.9 °C**; Full Red Color at **10.1 °C**.

Quality Control Verification: \_\_\_\_\_ **20160802**  
 Quality manager Date

Quality Assurance: Approved Lot # **S10070716**; \_\_\_\_\_ **2016-08-02**