

Safe-T-Vue® • Product No. 7201
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 10 Non-Reversible Temperature Indicator Product No. 7201

Lot # S10 030617

Expiration Date: 2019-03-06

Non - Chemistry

Mark ✓ for "Accepted" and X for "Not Accepted."

| Box | Adh. Label | Foil Lid | Close Function | Fill Level |
|--------------|-------------|--------------|-----------------|--------------------|
| ✓ Copy | ✓ Copy | ✓ Copy | ✓ Formed | ✓ Acceptable |
| ✓ Tack | ✓ Tack | ✓ Position | ✓ Locked/Stable | Underfill |
| ✓ Clean | ✓ Color | ✓ Clean Peel | Hinge | Overfill |
| ✓ Box Damage | Tray | Disk | ✓ Elasticity | Color Label |
| | ✓ Edges | ✓ Appearance | | ✓ Color |
| | ✓ Color | ✓ Fit | | |

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.

| Temperature Change and Corresponding Safe-T-Vue® Change | |
|--|--|
| Product Specification | Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C |
| Starting Test Temperatures | 1 to 4 °C |

| Test Parameter | AQL | Sample Size | Accept Number | Test Result | Pass/Fail |
|---------------------------------|-----|-------------|---------------|-------------|-----------|
| Full Color Response Temperature | 1% | 125 | 2 | PASS | PASS |

Quantity Approved: 59,250

Performance Data: Red Spots - Random 9.4 °C; Red Filling in White Center 9.8 °C; Full Red Color at 10.0 °C.

Quality Control Verification: [Signature] 2017-04-03
 Quality manager Date

Quality Assurance: Approved Lot # S10030617; [Signature] 2017-04-05