

Safe-T-Vue® • Product No. 7206
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 6 Non-Reversible Temperature Indicator Product No. **7206**

Lot # S6 051217

Expiration Date: **2019-05-12**

Non - Chemistry

Mark ✓ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
✓ Copy	✓ Copy	✓ Copy	✓ Formed	✓ Acceptable
✓ Tack	✓ Tack	✓ Position	✓ Locked/Stable	Underfill
✓ Clean	✓ Color	✓ Clean Peel	Hinge	Overfill
✓ Box Damage	Tray	Disk	✓ Elasticity	Color Label
	✓ Edges	✓ Appearance		✓ Color
	✓ Color	✓ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	50	0	Pass	Pass

Quantity Approved: _____ 7,000 _____

Performance Data: Red Spots - Random **5.2 °C**; Red Filling in White Center **5.6 °C**; Full Red Color at **6.0 °C**.

Quality Control Verification: _____ **2017-06-14**
 Quality manager Date

Quality Assurance: Approved Lot # **S6051217** ; _____ **2017-06-14**
 Quality manager Date