

Safe-T-Vue® • Product No. 7201
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® **10** Non-Reversible Temperature Indicator Product No. **7201**

Lot # S10 041817

Expiration Date: **2019-04-18**

Non - Chemistry

Mark for "Accepted" and X for "Not Accepted."

| Box | Adh. Label | Foil Lid | Close Function | Fill Level |
|--|---|--|---|--|
| <input checked="" type="checkbox"/> Copy | <input checked="" type="checkbox"/> Copy | <input checked="" type="checkbox"/> Copy | <input checked="" type="checkbox"/> Formed | <input checked="" type="checkbox"/> Acceptable |
| <input checked="" type="checkbox"/> Tack | <input checked="" type="checkbox"/> Tack | <input checked="" type="checkbox"/> Position | <input checked="" type="checkbox"/> Locked/Stable | <input type="checkbox"/> Underfill |
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Color | <input checked="" type="checkbox"/> Clean Peel | Hinge | <input type="checkbox"/> Overfill |
| <input checked="" type="checkbox"/> Box Damage | Tray | Disk | <input checked="" type="checkbox"/> Elasticity | Color Label |
| | <input checked="" type="checkbox"/> Edges | <input checked="" type="checkbox"/> Appearance | | <input checked="" type="checkbox"/> Color |
| | <input checked="" type="checkbox"/> Color | <input checked="" type="checkbox"/> Fit | | |

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.


| Temperature Change and Corresponding Safe-T-Vue® Change | |
|--|--|
| Product Specification | Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 8±0.3°, Safe-T-Vue 10±0.3°C |
| Starting Test Temperatures | 1 to 4 °C |

| Test Parameter | AQL | Sample Size | Accept Number | Test Result | Pass/Fail |
|---------------------------------|-----|-------------|---------------|-------------|-----------|
| Full Color Response Temperature | 1% | 125 | 2 | Pass | Pass |

Quantity Approved: 58,850

Performance Data: Red Spots - Random **9.6°C**; Red Filling in White Center **9.9°C**; Full Red Color at **10.1°C**.

Quality Control Verification:  2017-06-14
 Quality manager Date

Quality Assurance: Approved Lot # S10041817 :  2017-06-14
 Quality Manager Date