

Safe-T-Vue® • Product No. 7201-50 and 7201-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # S10 032918

Expiration Date: **2020-03-29**

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	" Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.williamlabs.com.

Temperature Change and Corresponding Safe-T-Vue [®] Change			
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C		
Starting Test Temperatures	1 to 4 °C		

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	Pass	Pass

Quantity Approved: <u>121,850</u>					
Performance Data: Red Spots - Random 9.5 ° C; Red Filling in White Center 9.8 ° C; Full Red Color at 10.1 ° C.					
Quality Control Verification: M SAccident SAccident Control Verification:					
Quality Assurance: Approved Lot # <u>\$\0037°\(\frac{8}{};</u>	Quality manager	2018-05-23 Date			

QCLF QAP0102-2 STV QA Lot Performance Report STV Rev. 15 Registry Doc. #2046